



ThermoWorks

CERTIFICATE OF CALIBRATION

Calibrated by ETI Ltd for:
ThermoWorks, Inc.
1762 W. 20 S., Unit 100
Lindon, UT 84042

Date of Issue: 25TH MARCH

Certificate Issued by: ST

Instrument Description: TempTest®

Instrument Serial Number: D14120827

Ambient Temperature: 72 °F ±4 °F Ambient Humidity: <80 %rh

Reference Instrument UKAS Certificate Number: CN153120 / N7110582T

Results: Test Temperature °F	Instrument Reading °F
32.0	32
212.0	212

Uncertainty of Calibration: ±0.7 °F ±1 digit

This is to certify that the above Instrument has been calibrated against laboratory Standards which are traceable, via International Agreement to all major National Standards, including NIST.

Authorised by:

Calibration Manager for ETI Ltd